



POTVRDA O TIPNOM ODOBRENJU
TYPE APPROVAL CERTIFICATE

Br / No **02-001500/023129**

Potvrđuje se da je ovaj proizvod, u skladu s Pravilima za klasifikaciju brodova, Dio 1. - Opći propisi, Odjeljak 3. - Tipno odobrenje proizvoda, odobren za ugradnju na brodove i postrojenja pod nadzorom Hrvatskog registra brodova.
This is to certify that this product on the basis of the Rules for the classification of ships, Part 1 - General requirements, Chapter 3 - Type approval of products, is approved for use on board ships and installations under supervision of the Croatian Register of Shipping.

TIP I OPIS PROIZVODA:

TYPE AND DESCRIPTION OF PRODUCT:

MANOMETAR **TIP M 820 OZNAKA "Y" 1, 2, 3 i 4**
PRESSURE GAUGE *TYPE M 820 DESIGNATION "Y" 1, 2, 3 and 4*

PROIZVOĐAČ:
MANUFACTURER:

INOL d.o.o.
Poljubinj 89 F, 5220 TOLMIN
SLOVENIJA • SLOVENIA

PROIZVOD UDOVOLJAVA SLIJEDEĆIM PRAVILIMA/PROPISIMA:
THE PRODUCT MEETS FOLLOWING RULES/REGULATIONS:

HRB: Pravila za klasifikaciju brodova, Dio 8. - Cjevovodi.
CRS: Rules for the classification of ships, Part 8. - Piping.

DETALJI O PROIZVODU I UVJETIMA CERTIFIKACIJE DANI SU NA POLEĐINI.
FURTHER DETAILS OF THE PRODUCT AND CONDITIONS FOR CERTIFICATION ARE GIVEN OVERLEAF.

ODOBRENJE VRIJEDI DO: 2014-12-18
APPROVAL IS VALID UNTIL:

Mjesto i datum Split, 2011-06-27
Place and date

M.P. _____
(Seal)

Goran Begović, dipl.ing.

Napomena: Ova potvrda nije valjana ako se proizvod ili način njegove proizvodnje promijenio u odnosu na tipno ispitani proizvod. Proizvođač treba obavijestiti Hrvatski registar brodova o bilo kojoj izmjeni na proizvodu kako bi ishodio valjanu potvrdu.
Note: This certificate is not valid for equipment, the design or manufacture of which has been varied or modified from the specimen tested. The manufacturer should notify Croatian Register of Shipping of any modification or changes to the product in order to obtain a valid certificate.

PODROBNIJI OPIS PROIZVODA:

DETAIL PRODUCT DESCRIPTION:

Manometar djeluje na principu bourdonove cijevi.

Pressure gauge utilizes bourdon tube effect.

PODRUČJE PRIMJENE / POSEBNI I GRANIČNI UVJETI:

APPLICATIONS / LIMITATIONS:

Mjerenje tlaka plinova i tekućina.

U skladu s nacrtom br. IM-820-02 kode broj 820 A Y-00-00 za "Y" 1, 2, 3 i 4

Razred točnosti manometra 1,0.

Pressure measurement of gases and liquids.

In accordance with drawing No. IM-820-02 code No. 820 A Y-00-00 for "Y" 1, 2, 3 and 4

Class of accuracy of pressure gauge 1,0.

DOKUMENTACIJA ZA TIPNO ODOBRENJE:

TYPE APPROVAL DOCUMENTATION:

Nacrt broj: IM-820-02 kode broj 820 AY-00-00

Drawing No: IM-820-02 code No. 820 AY-00-00

Nacrt broj: IM-820-02 kode broj 820-00-00

Drawing No: IM-820-02 code No. 820-00-00

Sastavnica br.: IM-820-02 kode broj 820-00-00

Material specification No.: IM-820-02 code No. 820-00-00

OZNAČAVANJE PROIZVODA:

MARKING OF PRODUCT:

- | | |
|----------------------|-----------------------|
| - oznaka proizvođača | - manufacturer's mark |
| - tip | - type |
| - serijski broj | - serial No. |
| - razred točnosti | - class |

UVJETI ZA CERTIFIKACIJU:

CONDITIONS OF CERTIFICATION:

Raščlana

08.20

Zamjenjuje potvrdu

02-001110/022187