



**POTVRDA O ODOBRENJU USLUŽNE TVRTKE**  
**CERTIFICATE OF APPROVAL OF THE SERVICE SUPPLIER**

Br / No **02-001997/023737**

Potvrđuje se da je ova uslužna tvrtka odobrena u skladu s Pravilima za klasifikaciju brodova, Dio 1. – Opći propisi, Odjeljak 4. – Odobrenje proizvođača i uslužnih tvrtki.

*This is to certify that this service supplier is approved on the basis of the Rules for the classification of ships, Part 1 – General Requirements, Chapter 4 – Approval of manufacturers and service suppliers.*

**TVRTKA:**

*FIRM:*

**ULJANIK – BRODOGRADILIŠTE d.d.**

*Primopredaja i kontrola*

**URED AUTOMATIKE**

*Flaciusova 1, 52100 PULA*

**HRVATSKA • CROATIA**

**PODRUČJE ODOBRENJA:**

*FIELD OF APPROVAL:*

**Ispitivanje i umjeravanje manometara, tlaka od 0 do 350 bar i razreda točnosti 0,4 i niže.**

**Ispitivanje i umjeravanje manometara, tlaka od 400 do 2500 bar i razreda točnosti 0,6 i niže.**

**Ispitivanje i umjeravanje termometara, temperature –35 do 600°C i razreda točnosti +/-0,2.**

**Ispitivanje i umjeravanje termoparova, temperature –270 do 1372°C i razreda točnosti +/-0,1.**

**Ispitivanje i umjeravanje vakuummetara, tlaka od 0 do –1,0 bar i razreda točnosti 0,4 i niže.**

**Ispitivanje sigurnosnih ventila.**

*Testing and gauging of manometers, pressure range 0 to 350 bar and accuracy class 0,4 and lower.*

*Testing and gauging of manometers, pressure range 400 to 2500 bar and accuracy class 0,6 and lower.*

*Testing and gauging of manometers, pressure range –35 to 600°C and accuracy class +/-0,2.*

*Testing and gauging of thermocouples, temperature range –270 to 1372°C and accuracy class +/-0,1.*

*Testing and gauging of vacuummeters, pressure range 0 to –1,0 bar and accuracy class 0,4 and lower.*

*Testing of safety valves.*

**ODOBRENJE VRIJEDI DO:**  
*APPROVAL IS VALID UNTIL:*

**2019-06-05**

**UZ UVJETE NAVEDENE NA POLEDINI**  
*SUBJECT TO CONDITION OVERLEAF*

**Mjesto i datum**  
*Place and date*

Split, 2015-05-07

**M.P.**  
*(Seal)*

**(Potpis – Signature)**

Marinko Popović, dipl.ing.

**Ispitivanje i umjeravanje manometara, termometara i termoparova odobreno je dopisom T/S/405-SG/ od 05.06.95.**

*Testing and gauging of manometers, thermometers and thermocouples is approved with letter T/S/405-SG/ dated 05.06.95.*

**Uslužna tvrtka se uvrštava u Popis odobrenih uslužnih tvrtki i odobrena je sve dotle dok se pridržava odredaba gore navedenih Pravila.**  
*The service supplier is enlisted in the List of approved services suppliers and is approved until the provisions of a.m. Rules are followed.*

**Raščlana**

60.20

**Zamjenjuje potvrdu**

02-001494/023098